



广微测
Gmicro Testing



201819000883



中国认可
国际互认
检测
TESTING
CNAS L1747

广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

分析检测报告

REPORT FOR ANALYSIS

报告编号

Report No.

2024SP05987R02

样品名称

Name of Sample

智能紫外线消毒装置

委托单位

Applicant

苏州迅奇云消卫生与环境技术有限公司

检测类型

Test Type

委托检测

单位地址: 广州市先烈中路 100 号大院 66 号楼

Address:

Building 66, No.100 Central Xian Li Road, Guangzhou, China

邮政编码: 510070

Postcode:

电话号码: (020)87137666

Tel:

传真号码: (020)87137668

Fax:

网 址: www.gddcm.com

Website:



广微测
Gmicro Testing

广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

分析检测报告

REPORT FOR ANALYSIS



报告编号 (Report No.) 2024SP05987R02

校验码 (Verification Code): 6PQXLH6P

样品名称 Name of Sample	智能紫外线消毒装置	检测类型 Test Type	委托检测
委托单位 Applicant	苏州迅奇云消卫生与环境技术有限公司	地址 Address	江苏省昆山市开发区 杜鹃路 555 号 C2
样品来源 Sample Source	委托方送检	样品数量 Sample Quantity	1 台
样品规格和批号 Spec and Lot No of Sample	UVhero S5	样品状态和特性 State and Characteristic	机器
接样日期 Sample Received Date	2024-05-22	验讫日期 Completion Date	2024-06-06
检测项目 Item Tested	臭氧泄漏量		
检测依据和方法 Test Standard and Method	GB 28232-2020 附录 A.2.1		
检测结论 Test Conclusion	送检样品所检项目的实测数据见本检测报告续页。		
备注 Remarks	生产厂家: 苏州迅奇云消卫生与环境技术有限公司, 地址: 江苏省昆山市开发区杜鹃路 555 号 C2 (由委托方提供)		

签发日期: 2024-06-06

Issue Date

(机构盖章 Official Seal)

制表:
Editor

李韵倩

审核:
Verifier

陈维

批准:
Approver

尹



广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

分析检测结果

ANALYSIS AND TEST RESULT

报告编号 (Report No.) : 2024SP05987R02

臭氧泄漏量

1. 方法简述: 将送检样品置于 30 m³ 空间中央运行样品, 测点高度 1.5 m, 每 5 min 记录一次, 至 60 min 时结束。
2. 检测结果:

开启时间	检测结果	单位	检测方法
0 min	0.00	mg/m ³	GB 28232-2020 附录 A.2.1
5 min	0.02	mg/m ³	
10 min	0.01	mg/m ³	
15 min	0.01	mg/m ³	
20 min	0.01	mg/m ³	
25 min	0.02	mg/m ³	
30 min	0.01	mg/m ³	
35 min	0.00	mg/m ³	
40 min	0.00	mg/m ³	
45 min	0.01	mg/m ³	
50 min	0.01	mg/m ³	
55 min	0.01	mg/m ³	
60 min	0.01	mg/m ³	
平均值	<0.01	mg/m ³	

(接下页)



广微测
Gmicro Testing

广东省微生物分析检测中心

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

分析检测结果

ANALYSIS AND TEST RESULT

报告编号 (Report №.) : 2024SP05987R02

(接上页)

样品照片:



(接下页)

报告编号 (Report №.) : 2024SP05987R02

注意事项

Notice Items

1. 检测报告无本机构检验检测专用章、骑缝章无效。

The test report is invalid without the special stamp of testing and paging seal.



广 微 测

Gmicro Testing

2. 检测报告无制表、审核、批准人签字无效。

The test report is invalid without the signatures of the editor, the verifier and the approver.

3. 检测报告涂改增删无效。

The test report is invalid if being supplemented, deleted or altered.

4. 未经本机构书面同意，不得部分复制（全部复制除外）检测报告。

Without prior written permission, the test report cannot be reproduced, except in full.

5. 除非另有说明，检测报告仅对来样负责。

Unless otherwise stated, the results shown in this test report refer only to the sample(s) tested.

6. 对检测报告有异议的，应于收到报告之日起十五日内提出，逾期不予受理。

Any dispute of the test report must be raised to the testing institution within 15 days after the test report is received, exceeding which the dispute will not be accepted.

7. 对送检样品，样品信息由委托方提供，本机构不对其真实性负责。

For the sample(s) submitted for testing, the sample information is provided by the applicant, and this testing institution is not responsible for its authenticity.

